

**Ellsworth Community College Metrology Laboratory
Calibration Report**

Ellsworth Community College
1100 College Ave.
Iowa Falls, IA 50126
ph.: 641.648.8737
fax: 641.648.3128

TF-0402
04.01.2010



ELLSWORTH
COMMUNITY COLLEGE

Receiving Date: 07.19.2011
Test Date: 07.20.2011
Report Date: 07.23.2011

State Test No.: IA1.11.0035
Set Serial No.: na
Accreditation: NIST Recognition April 25th 2011 - Dec. 31st 2011

Hawkeye State Scale
5040 Blair Forest Way
NE Ste F
Cedar Rapids IA 52404

Contact: John Fischbeck
Ph.: 319.364.4173
P.O. No.: na
SOP: 8 Modified Substitution
Tech.: A.B.

Item(s) Submitted: CI Standards
Manufacturer: na
Class: F
Equip. ID#: na
Condition: Good
Temp. C: 22
Pressure mmHg: na
% Relative Humidity: 60

Correction (g)

Nominal Value	Unit	Serial/ID No.	As Found (g)	As Left (g)	Density	Tol.(g)	UNC (g) (k=2)
50	lb	4-26	-1.26	-0.126	7	2.3	0.27
50	lb	4-35	-4.63	0.2	7	2.3	0.27
50	lb	4-31	-5.67	0.1	7	2.3	0.27
50	lb	4-32	1.45	1.45	7	2.3	0.27

The resulting tolerance class of the weight is determined by combining the correction of the weight and the uncertainty of the measurement. The corrections given above correlate to a conventional mass scale versus 8.0 g/cm³ density of 1.2 mg/cm³. The items listed above have been compared to Standards of Iowa which are currently in control. These standards are traceable to the National Institute of Standards and Technology (NIST). Calibration processes were monitored and found to be in control. Uncertainty calculations conform to NIST Note 1297. Results apply to item identified in this report only.

Ellsworth Community College Metrology Laboratory

Laboratory Metrologist
Andrew Blackburn

Reviewed By
Ellsworth Community College Metrology Secretary