



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection

Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate for calibration work performed for: HAWKEYE STATE SCALE, INC.

1357 HWY 965 NW
SWISHER, IA 52338
(563) 554-1406

Date Received: January 09, 2025
Date of Calibration: January 09, 2025
Date Issued: January 13, 2025

State Test No.: W25-006

Uncertainty Statement

For the mass standards used in this calibration, some uncertainty components were assessed through a Type A evaluation, the method for assessing uncertainty by a statistical analysis of measured quantity values obtained under defined measurement conditions. In addition, other components were assessed from a Type B evaluation of standard uncertainty, based on scientific judgement using all of the relevant information available. The combined standard uncertainty was multiplied by a statistically determined coverage factor to provide an expanded uncertainty. The expanded uncertainty defines an interval having a level of confidence of approximately 95 percent, assuming normal distribution. The expanded uncertainty presented in this report is consistent with the ISO/IEC Guide to the Expression of Uncertainty in Measurement using the Root Sum Squares method (JCGM 100:2008).

Traceability Statement

The standards used by the Wisconsin State laboratory demonstrate an unbroken traceable chain to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory maintains documented calibration intervals and uses documented procedures, all under the performance of trained personnel who demonstrate suitable measurement assurance for the information listed in this calibration certificate. The laboratory test number identified above is the unique test number to be used in referencing measurement traceability for the artifacts identified in this certificate. The State Standards are traceable to the SI unit for mass, the kilogram.

Conformity Statement


These results relate only to the items calibrated in this certificate. Field standards and weight carts are calibrated based on guidance described in NIST Handbook 105-1 (2019) and NIST Handbook 105-8 (2019), respectively, using NISTIR 6969: Selected Laboratory Measurement Practices and Procedures to Support Basic Mass Calibrations (2019). Field standards calibrated to NIST Class F, ASTM 5, and ASTM 6 tolerances are usable for testing class III, III L, and IIII weighing devices, following NIST Handbook 44, Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices. Field standards calibrated to NIST Class F, ASTM 5, or ASTM 6 tolerances are not suitable for testing class I and class II weighing devices, which must be tested with field standards of higher precision than NIST Class F, ASTM 5, or ASTM 6. Weights calibrated to ASTM 7 tolerances by this laboratory cannot be used for testing commercial weighing devices. Field standards calibrated to ASTM Standard Specification E617-23 are not checked for density [Stainless steel weights are assumed 8.0 grams per cubic centimeter], or for magnetism.

Decision Rule

All calibrated weights and weight carts that are determined to have a mass correction such that: $| \text{Correction} | > (\text{Tolerance} - \text{Uncertainty})$ are considered to have failed to meet the applicable tolerance. It is the decision rule of the Wisconsin State laboratory that all calibrated weights and weight carts that are determined to have a mass correction such that: $| \text{Correction} | > (0.95 \times \text{Tolerance} - \text{Uncertainty})$ will be adjusted to be closer to zero mass correction, even if the mass correction of the weights and weight carts originally met the applicable tolerance. Customers may request exceptions to this decision rule.

The following standard(s) were used: 1000 lb: 392


This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.



Paul Masterson, Lead Metrologist

Page 1 of 2

3601 Galleon Run • Madison, WI 53718 • (608) 224-4910



Justin Lien, Laboratory Director



State of Wisconsin
Governor Tony Evers

Department of Agriculture, Trade and Consumer Protection

Secretary Randy Romanski

Wisconsin Weights and Measures Laboratory

Calibration Certificate

Date Received: January 09, 2025

State Test No.: W25-006

Date of Calibration: January 09, 2025

Item(s) Submitted: Cast Weight

Date Issued: January 13, 2025

Manufacturer: Various

Condition: Good, Acceptable for Calibration

Tolerance Class: NIST HB 105-1 (1990), Class F

Customer: HAWKEYE STATE SCALE, INC.

Balance ID#: 10

Address: 1357 HWY 965 NW

Procedure Used: NISTIR 6969 (2019), SOP 8

SWISHER, IA 52338

Temperature: 20.7 °C

Contact: ANTHONY KRUSE

Relative Humidity: 42.0 %


Phone: (563) 554-1406

Pressure: 739.5 mmHg

| Nominal Mass | Mass Unit | Serial No. | Conventional Mass Correction (mg) As Found | Conventional Mass Correction (mg) As Left | NIST HB 105-1 (1990), Class F As Found | NIST HB 105-1 (1990), Class F As Left | Uncertainty (mg) | Coverage Factor (k) |
|--------------|-----------|------------|---|--|---|--|------------------|---------------------|
| 1000 | lb | 2-19 | -29,000 | -29,000 | Pass | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-12 | -39,800 | 20,600 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-11 | -57,300 | 22,100 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-17 | -96,200 | 18,500 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-20 | -132,800 | 19,000 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-18 | -76,100 | 19,700 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-8 | -58,000 | 20,100 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-3 | -33,800 | -33,800 | Pass | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-4 | -39,800 | 17,700 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-2 | -54,800 | 24,100 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-10 | -53,500 | 17,800 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-1 | -32,500 | -32,500 | Pass | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-6 | -45,400 | 15,800 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-5 | -54,800 | 16,200 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-9 | -41,600 | 18,500 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-16 | -65,800 | -1,400 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-14 | -71,700 | 16,200 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-13 | -70,900 | 15,000 | Fail | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-15 | -21,100 | -21,100 | Pass | Pass | 5,900 | 2.00 |
| 1000 | lb | 2-7 | -33,500 | -33,500 | Pass | Pass | 5,900 | 2.00 |

The following standard(s) were used: 1000 lb: 392

This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.



Paul Masterson, Lead Metrologist

Page 2 of 2

3601 Galleon Run • Madison, WI 53718 • (608) 224-4910



Justin Lien, Laboratory Director

Traceable Certificate Number: 3996903A

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2E5

Comments:

Key Notes

| | |
|--------------|---|
| Finish | Indicates the weight does not meet the finish requirements |
| Material | Indicates the weight does not meet the material requirements |
| New Wt | Indicates new weight |
| Missing Wt | Indicates replaced missing weight with new weight |
| Damaged Wt | Indicates replaced damaged weight |
| Replaced OOT | Indicates replaced out of tolerance weight |
| OOT | Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | Indicates replaced magnetic weight |
| Design | Indicates the weight does not meet the design or shape requirements |
| Repainted | Indicates the weight was repainted after As Found obtained |
| Other | See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.



Cleaning Levels

- A Dusted with brush or cloth
- B Spot cleaned with ethyl alcohol
- C Full surface cleaned with ethyl alcohol
- D Spot cleaned with non-alcohol solvent followed by ethyl alcohol
- E Full surface cleaned with non-alcohol solvent followed by ethyl alcohol
- F No cleaning performed

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |

Austin Anderson, Lab Technician



Prepared By:

Rice Lake Weighing Systems® • PN 64784 • 6/25

230 West Coleman Street • Rice Lake, WI 54686 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Traceable Certificate Number: 3996903A

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

Temperature Range: 20.15 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 43 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb C2E5 | | 1000.021 | 9608 | 1000.005 | 2047 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903B

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2E6

Comments:

Key Notes

| | |
|--------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements |
| Material | ✖ Indicates the weight does not meet the material requirements |
| New Wt | ✳ Indicates new weight |
| Missing Wt | ✳ Indicates replaced missing weight with new weight |
| Damaged Wt | ✳ Indicates replaced damaged weight |
| Replaced OOT | ✳ Indicates replaced out of tolerance weight |
| OOT | ✖ Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | ★★ Indicates replaced magnetic weight |
| Design | ✖ Indicates the weight does not meet the design or shape requirements |
| Repainted | █ Indicates the weight was repainted after As Found obtained |
| Other | ✚ See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.



Cleaning Levels

| | |
|---|---|
| A | Dusted with brush or cloth |
| B | Spot cleaned with ethyl alcohol |
| C | Full surface cleaned with ethyl alcohol |
| D | Spot cleaned with non-alcohol solvent followed by ethyl alcohol |
| E | Full surface cleaned with non-alcohol solvent followed by ethyl alcohol |
| F | No cleaning performed |

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |


Austin Anderson, Lab Technician



Prepared By:

Rice Lake Weighing Systems® • P/N 64784-06/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Traceable Certificate Number: 3996903B

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

Temperature Range: 20.15 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 43 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb | C2E6 | 1000.032 | 14609 | 1000.016 | 7047 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903C

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2E7


Comments:

Key Notes

| | |
|--------------|---|
| Finish | Indicates the weight does not meet the finish requirements |
| Material | Indicates the weight does not meet the material requirements |
| New Wt | Indicates new weight |
| Missing Wt | Indicates replaced missing weight with new weight |
| Damaged Wt | Indicates replaced damaged weight |
| Replaced OOT | Indicates replaced out of tolerance weight |
| OOT | Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | Indicates replaced magnetic weight |
| Design | Indicates the weight does not meet the design or shape requirements |
| Repainted | Indicates the weight was repainted after As Found obtained |
| Other | See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. **This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.**



Cleaning Levels

| | |
|---|---|
| A | Dusted with brush or cloth |
| B | Spot cleaned with ethyl alcohol |
| C | Full surface cleaned with ethyl alcohol |
| D | Spot cleaned with non-alcohol solvent followed by ethyl alcohol |
| E | Full surface cleaned with non-alcohol solvent followed by ethyl alcohol |
| F | No cleaning performed |

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |

Austin Anderson, Lab Technician



Prepared By:

Rice Lake Weighing Systems® • PN 64784 • 6/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

Traceable Certificate Number: 3996903C

Client: HAWKEYE STATE SCALE
Date Calibrated: 26 Jan 2026

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Temperature Range: 20.15 °C
Pressure Range: 727.69 mmHg
Relative Humidity Range: 43 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb C2E7 | | 1000.032 | 14609 | 1000.016 | 7047 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903D

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2E8


Comments:

Key Notes

| | |
|--------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements |
| Material | ✖ Indicates the weight does not meet the material requirements |
| New Wt | ❖ Indicates new weight |
| Missing Wt | ▲ Indicates replaced missing weight with new weight |
| Damaged Wt | ✖ Indicates replaced damaged weight |
| Replaced OOT | ★ Indicates replaced out of tolerance weight |
| OOT | ☒ Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | ★★ Indicates replaced magnetic weight |
| Design | ☒ Indicates the weight does not meet the design or shape requirements |
| Repainted | ■ Indicates the weight was repainted after As Found obtained |
| Other | ✚ See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.



Austin Anderson

Austin Anderson, Lab Technician



Prepared By:

Rice Lake Weighing Systems® • PN 04784 • 6/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

Traceable Certificate Number: 3996903D

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Temperature Range: 20.15 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 43 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb | C2E8 | 1000.036 | 16109 | 1000.019 | 8548 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903E

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2E9

Comments:

Key Notes

| | |
|--------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements |
| Material | ✖ Indicates the weight does not meet the material requirements |
| New Wt | ❖ Indicates new weight |
| Missing Wt | ▲ Indicates replaced missing weight with new weight |
| Damaged Wt | ✖ Indicates replaced damaged weight |
| Replaced OOT | ★ Indicates replaced out of tolerance weight |
| OOT | ✖ Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | ★★ Indicates replaced magnetic weight |
| Design | ✖ Indicates the weight does not meet the design or shape requirements |
| Repainted | ■ Indicates the weight was repainted after As Found obtained |
| Other | ✚ See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. **This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.**



| Key Notes | | Cleaning Levels | |
|--------------|---|-----------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements | A | Dusted with brush or cloth |
| Material | ✖ Indicates the weight does not meet the material requirements | B | Spot cleaned with ethyl alcohol |
| New Wt | ❖ Indicates new weight | C | Full surface cleaned with ethyl alcohol |
| Missing Wt | ▲ Indicates replaced missing weight with new weight | D | Spot cleaned with non-alcohol solvent followed by ethyl alcohol |
| Damaged Wt | ✖ Indicates replaced damaged weight | E | Full surface cleaned with non-alcohol solvent followed by ethyl alcohol |
| Replaced OOT | ★ Indicates replaced out of tolerance weight | F | No cleaning performed |
| OOT | ✖ Indicates correction plus or minus Uncertainty greater than or equal to MPE | | |
| Magnetic Wt | ★★ Indicates replaced magnetic weight | | |
| Design | ✖ Indicates the weight does not meet the design or shape requirements | | |
| Repainted | ■ Indicates the weight was repainted after As Found obtained | | |
| Other | ✚ See comments above | | |

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |


Austin Anderson, Lab Technician



Prepared By:

Rice Lake Weighing Systems® • PN 64784 • 6/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

Traceable Certificate Number: 3996903E

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Temperature Range: 20.15 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 43 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb | C2E9 | 1000.024 | 11108 | 1000.008 | 3547 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903F

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2EA

Comments:

Key Notes

| | |
|--------------|---|
| Finish | Indicates the weight does not meet the finish requirements |
| Material | Indicates the weight does not meet the material requirements |
| New Wt | Indicates new weight |
| Missing Wt | Indicates replaced missing weight with new weight |
| Damaged Wt | Indicates replaced damaged weight |
| Replaced OOT | Indicates replaced out of tolerance weight |
| OOT | Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | Indicates replaced magnetic weight |
| Design | Indicates the weight does not meet the design or shape requirements |
| Repainted | Indicates the weight was repainted after As Found obtained |
| Other | See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. **This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.**



Cleaning Levels

- A Dusted with brush or cloth
- B Spot cleaned with ethyl alcohol
- C Full surface cleaned with ethyl alcohol
- D Spot cleaned with non-alcohol solvent followed by ethyl alcohol
- E Full surface cleaned with non-alcohol solvent followed by ethyl alcohol
- F No cleaning performed

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |



Austin Anderson, Lab Technician


Prepared By:

Rice Lake Weighing Systems® • PN 64784 • 6/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

Traceable Certificate Number: 3996903F

Client: HAWKEYE STATE SCALE
Date Calibrated: 26 Jan 2026

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Temperature Range: 20.13 °C
Pressure Range: 727.69 mmHg
Relative Humidity Range: 44 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb C2EA | | 1000.033 | 15109 | 1000.017 | 7547 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903G

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593
Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025
Date Calibrated: 26 Jan 2026
Recalibration Date: No Recall Requested
NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28
Procedure: WI05-0023 Rev. L
Condition of Weights: New
Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2EB

Comments:

Key Notes

| | |
|--------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements |
| Material | ✖ Indicates the weight does not meet the material requirements |
| New Wt | ◆ Indicates new weight |
| Missing Wt | ▲ Indicates replaced missing weight with new weight |
| Damaged Wt | ✖ Indicates replaced damaged weight |
| Replaced OOT | ★ Indicates replaced out of tolerance weight |
| OOT | ☒ Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | ★★ Indicates replaced magnetic weight |
| Design | ☒ Indicates the weight does not meet the design or shape requirements |
| Repainted | ■ Indicates the weight was repainted after As Found obtained |
| Other | ✚ See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.



| Key Notes | | Cleaning Levels | |
|--------------|---|-----------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements | A | Dusted with brush or cloth |
| Material | ✖ Indicates the weight does not meet the material requirements | B | Spot cleaned with ethyl alcohol |
| New Wt | ◆ Indicates new weight | C | Full surface cleaned with ethyl alcohol |
| Missing Wt | ▲ Indicates replaced missing weight with new weight | D | Spot cleaned with non-alcohol solvent followed by ethyl alcohol |
| Damaged Wt | ✖ Indicates replaced damaged weight | E | Full surface cleaned with non-alcohol solvent followed by ethyl alcohol |
| Replaced OOT | ★ Indicates replaced out of tolerance weight | F | No cleaning performed |
| OOT | ☒ Indicates correction plus or minus Uncertainty greater than or equal to MPE | | |
| Magnetic Wt | ★★ Indicates replaced magnetic weight | | |
| Design | ☒ Indicates the weight does not meet the design or shape requirements | | |
| Repainted | ■ Indicates the weight was repainted after As Found obtained | | |
| Other | ✚ See comments above | | |

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |

Austin Anderson
Austin Anderson, Lab Technician



Prepared By:

Rice Lake Weighing Systems® • PN 64784-06/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

Traceable Certificate Number: 3996903G

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Temperature Range: 20.13 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 44 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb | C2EB | 1000.028 | 12608 | 1000.011 | 5047 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903H

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2EC

Comments:

Key Notes

| | |
|--------------|---|
| Finish | Indicates the weight does not meet the finish requirements |
| Material | Indicates the weight does not meet the material requirements |
| New Wt | Indicates new weight |
| Missing Wt | Indicates replaced missing weight with new weight |
| Damaged Wt | Indicates replaced damaged weight |
| Replaced OOT | Indicates replaced out of tolerance weight |
| OOT | Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | Indicates replaced magnetic weight |
| Design | Indicates the weight does not meet the design or shape requirements |
| Repainted | Indicates the weight was repainted after As Found obtained |
| Other | See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. **This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.**

Austin Anderson, Lab Technician

NVLAP
CALIBRATION
NVLAP LAB CODE 105001-0

ILAC-MRA

Prepared By:

Rice Lake Weighing Systems® • PN 64784 • 6/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:

ACCREDITED
Certificate #4363.01



Certificate of Weight Calibration

Traceable Certificate Number: 3996903H

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Temperature Range: 20.13 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 44 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb C2EC | | 1000.017 | 7607 | 1000.000 | 46 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 39969031

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2ED

Comments:

Key Notes

| | |
|--------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements |
| Material | ✖ Indicates the weight does not meet the material requirements |
| New Wt | ◆ Indicates new weight |
| Missing Wt | ▲ Indicates replaced missing weight with new weight |
| Damaged Wt | ☒ Indicates replaced damaged weight |
| Replaced OOT | ★ Indicates replaced out of tolerance weight |
| OOT | ☒ Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | ★★ Indicates replaced magnetic weight |
| Design | ☒ Indicates the weight does not meet the design or shape requirements |
| Repainted | █ Indicates the weight was repainted after As Found obtained |
| Other | ✚ See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.




Cleaning Levels

- A Dusted with brush or cloth
- B Spot cleaned with ethyl alcohol
- C Full surface cleaned with ethyl alcohol
- D Spot cleaned with non-alcohol solvent followed by ethyl alcohol
- E Full surface cleaned with non-alcohol solvent followed by ethyl alcohol
- F No cleaning performed

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |



Austin Anderson, Lab Technician


Prepared By:

Rice Lake Weighing Systems® • PN 64784 • 6/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Traceable Certificate Number: 39969031

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

Temperature Range: 20.13 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 44 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb C2ED | | 1000.027 | 12108 | 1000.010 | 4547 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |

Traceable Certificate Number: 3996903J

Contractor: HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Purchase Order Number: 11593

Client: HAWKEYE STATE SCALE
1357 HWY 965 NW
info@hawkeyestatescale.com
SWISHER, IA 52338

Date Received: 23 Dec 2025

Date Calibrated: 26 Jan 2026

Recalibration Date: No Recall Requested

NIST Certificate Number: 684/O-0000046697 & 684/O-0000035150-22

If there are two NIST numbers, one or both may apply

Calibrated By: 28

Procedure: WI05-0023 Rev. L

Condition of Weights: New

Description of Weights: 1000 lb Cast Iron Heavy Capacity Weight, ASTM Class 6, S/N C2EE


Comments:

Key Notes

| | |
|--------------|---|
| Finish | ✖ Indicates the weight does not meet the finish requirements |
| Material | ✖ Indicates the weight does not meet the material requirements |
| New Wt | ❖ Indicates new weight |
| Missing Wt | ▲ Indicates replaced missing weight with new weight |
| Damaged Wt | ☒ Indicates replaced damaged weight |
| Replaced OOT | ★ Indicates replaced out of tolerance weight |
| OOT | ☒ Indicates correction plus or minus Uncertainty greater than or equal to MPE |
| Magnetic Wt | ★★ Indicates replaced magnetic weight |
| Design | ☒ Indicates the weight does not meet the design or shape requirements |
| Repainted | ■ Indicates the weight was repainted after As Found obtained |
| Other | ✚ See comments above |

Check with your local state agency for certification of compliance on Legal-for-Trade items. The weight accuracy class is referenced in the Description of Weights. Unless otherwise noted, the weights calibrated meet the requirements of the accuracy class. Results relate only to weights calibrated. The Surface Finishes of weights are evaluated visually. Weights are screened for magnetism using work instruction WI05-0035 when they are new, when requested by the customer or when weights are suspected of not meeting specifications. Density if measured is measured using OIML R111-1 (2004) method A2. Conventional Mass is reported based on a reference density of 8.0 g/cm³. The Uncertainty of Measurement is included in the determination of Maximum Permissible Error (MPE) Pass/Fail Criteria. The specifications for Maximum Permissible Error (MPE) can be found in NIST Handbook 105-1 (2019), NIST Handbook 105-1 (1990), ASTM E617-23 or OIML R111-1 (2004), manufacturer specifications or customer specifications.

The Uncertainty assigned to the Conventional Mass values are the result of the root-sum-square of the type A and type B components, calculated in accordance with NIST SOP 29 and the Guide to the expression of uncertainty in measurement, with coverage factor ($k=2$), to express the expanded uncertainty with an approximate 95.45% confidence level. This report is not to be used to claim product certification, approval, or endorsement by NVLAP, NIST, A2LA, or any government agency. This document and all data within, shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems.



Cleaning Levels

| | |
|---|---|
| A | Dusted with brush or cloth |
| B | Spot cleaned with ethyl alcohol |
| C | Full surface cleaned with ethyl alcohol |
| D | Spot cleaned with non-alcohol solvent followed by ethyl alcohol |
| E | Full surface cleaned with non-alcohol solvent followed by ethyl alcohol |
| F | No cleaning performed |

Material Abbreviations

| | | | |
|----|-----------------|----|---------------|
| AL | Aluminum | TA | Tantalum |
| SS | Stainless Steel | BR | Brass |
| CI | Cast Iron | PL | Platinum |
| IR | Iron | NS | Nickel Silver |
| MS | Mild Steel | OR | Other/Unknown |

Austin Anderson, Lab Technician



Prepared By:

Rice Lake Weighing Systems® • P/N 64784 • 6/25

230 West Coleman Street • Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967

Definitions: http://certs.ricelake.com/certs/0354_Term_Cert_Weight_Cal_Rev1.pdf

Page 1 of 2

26 Jan 2026

Issued Date:





Certificate of Weight Calibration

ISO/IEC 17025:2017 & ANSI/NCSL-Z540-1-1994 ACCREDITED

Traceable Certificate Number: 3996903J

Client: HAWKEYE STATE SCALE

Date Calibrated: 26 Jan 2026

Temperature Range: 20.13 °C

Pressure Range: 727.69 mmHg

Relative Humidity Range: 44 %

As Left Data (As Found Data is undifferentiated from As Left Data unless listed in As Found Data table)

| Nominal Value | Unique ID | True Mass (Same UOM as Nom.) | True Mass Corr. (mg) | Conv. Mass (Same UOM as Nom.) | Conv. Mass Corr. (mg) | (k=2) Unc. (± mg) | MPE (± mg) | MPE Pass (Y=Pass N=Fail) | Assumed Density (g/cm³) | Assumed Material | Const. Type | Balance Used | Reference Standard Set Used | Air Density (mg/cm³) | Clean Level |
|---------------|-----------|------------------------------|----------------------|-------------------------------|-----------------------|-------------------|------------|--------------------------|-------------------------|------------------|-------------|--------------|-----------------------------|----------------------|-------------|
| 1000 lb C2EE | | 1000.032 | 14609 | 1000.016 | 7047 | 8500 | 45000 | Y | 7.20 | CI | II | 851Q | 1095Q | 1.1482 | A |