



DEPARTMENT OF COMMERCE
WEIGHTS & MEASURES DIVISION

14305 Southcross Drive W #150
Burnsville, MN 55306-7008
mn.gov/commerce/
651.539.1555 FAX 952.435.4040
An equal opportunity employer

Receipt Date: March 31, 2025
Calibration Date: April 1, 2025
Certificate Date: April 1, 2025

Certificate No.: 403470-2
Set Serial No.: 1 to 9
Barcode: 204419

Calibration Certificate

HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338

Contact: ANGIE WELCHER
Phone: 319-364-4173
PO Number: None
Procedure: NIST SOP 8B (2019)
Technician ID: 23

Item(s) Submitted: 8 oz Fuel Sub Weights
Manufacturer: Rice Lake
Weight Type: II
Equipment ID: None
Condition: Acceptable
Temperature: 20.2 °C
Pressure: 737.5 mmHg
Relative Humidity: 47.6 %

Nominal Value		Serial No.	CM Correction (mg)		ASTM E617 (2018) Class		k	U (mg)
			As Received	As Left	As Received	As Left		
8 oz		1	16.76	16.76	6	6	2.03	0.54
8 oz		2	21.68	21.68	6	6	2.03	0.54
8 oz		3	25.44	5.23	*	5	2.03	0.54
8 oz		4	11.63	11.63	5	5	2.03	0.54
8 oz		5	11.55	11.55	5	5	2.03	0.54
8 oz		6	8.04	8.04	5	5	2.03	0.54
8 oz		7	18.08	18.08	6	6	2.03	0.54
8 oz		8	21.33	21.33	6	6	2.03	0.54
8 oz		9	21.56	21.56	6	6	2.03	0.54

Weight(s) as received (*) exceed ASTM class 6 tolerance.

Artifact conformance to ASTM E617 (2018) specifications of shape, material, markings, and type were evaluated. Tolerances were evaluated using ASTM E617 (2018) and MN SAP 20 (2020), which combines the conventional mass (CM) correction of the weight and the uncertainty of the measurement to evaluate the class. No other specifications were evaluated. The above CM corrections correspond to the mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. Uncertainty calculations contain the components in NIST SOP 8 (2019) combined using the root-sum-square method and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Calibration of items listed above used State of Minnesota Standards, which are currently in control. These standards are traceable to the SI through NIST. Calibration processes are monitored and in control at the time of calibration. Densities reported above are assumed unless noted. This calibration certificate shall not be reproduced, except in full, without written approval from the state of MN metrology laboratory, and the results only apply to items identified on this certificate.

Eric Johnson

Metrologist

Reviewed by:

Anna Pierce

Quality Manager, Signatory