

Receipt Date: April 20, 2026
Cal. Date: April 20, 2026
Certificate Date: April 22, 2026

Certificate No.: 404134-1
Set Serial No.: 040119K
Barcode: 204418

Calibration Certificate

HAWKEYE STATE SCALE
1357 HWY 965 NW
SWISHER, IA 52338
Contact: RENEE BISHOP
Phone: 319-364-4173
PO Number: None
Procedure: NIST SOP 33 (2019)
Technician ID: 26

Item(s) Submitted: 3000 lb Weight Cart
Manufacturer: KANAWHA
Cart Type: Gasoline
Equipment ID: None
Condition: ACCEPTABLE
Temperature: 20.6 °C
Pressure: 738.6 mmHg
Relative Humidity: 47.5 %

Nominal Value	Serial No.	CM Correction (g)		NIST HB 105-8 Class		k	U (g)
		As Received	As Left	As Received	As Left		
3000 lb	040119K	180	180	Meets	Meets	2.03	61

Artifact conformance to NIST HB105-8 (2019) specifications were evaluated except for tank cross sectional area, tank volume, tire tread pattern, filtration systems, sealed wheel bearings, functional remote control, or scale plate lengths and widths. Tolerances were evaluated using NIST HB 105-8 (2019) and MN SAP 20 (2025), which combines the conventional mass (CM) correction of the weight and the uncertainty of the measurement to evaluate the class. No other specifications were evaluated. The above CM corrections correspond to the mass scale versus 8.0 g/cm³ density and an air density of 1.2 mg/cm³ at 20 °C. Uncertainty calculations contain the components in NIST SOP 33 (2019) and conform to the ISO/IEC Guide to the Expression of Uncertainty in Measurement (2008), including coverage factors (*k*) calculated at the approximate 95.45 % confidence level. Calibration of items listed above used State of Minnesota Standards, which are currently in control. These standards are traceable to the SI through NIST. Calibration processes are monitored and in control at the time of calibration. This calibration certificate shall not be reproduced, except in full, without written approval from the state of MN metrology laboratory, and the results only apply to items identified on this certificate.

Calvin Crouch



Metrologist

Issued Date:

APR 23 2026

Page 1 of 1

Reviewed by:

Eric Johnson



Metrologist, Signatory



Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 105003-0. This certificate may not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the U.S. Government, and may not be reproduced, except in full, without written approval from the laboratory.